



Qualitätssicherung. Vom Landwirt bis zur Ladentheke.



QS Laboratory Performance Assessment Spring 2020

rating system laboratory performance assessment Spring 2020

Passed

all 8 active substances are identified correctly

and

at least 6 active substances are quantified with a satisfactory performance
(70-120% of the spiked level)

and

no additional active substance is identified

Failed *

not all of 8 active substances are identified

or

less than 6 active substances are quantified with a satisfactory performance
(70-120% of the spiked level)

or

one or more additional active substances are identified

* not successful, comments to the results of the performance assessment and the measures which have been taken have to be submitted to QS.